



03462-253443

E-mail: -dtowbbrb@rntcp.org
rntcpbirbhumb@gmail.com

GOVERNMENT OF WEST BENGAL
OFFICE OF THE CHIEF MEDICAL OFFICER OF HEALTH
DISTRICT TUBERCULOSIS WING
SURI, BIRBHUM

Memo No.: DTO /Birbhumb/76

Date:02/06/2009

Tender Notice

Sealed tenders by Registered post are invited from interested persons/firms/companies or organization for supply of Laboratory consumables (**Anx-I**) to District Health & Family Welfare Samity (RNTCP), Birbhumb.


Bid documents will be available from the office of the D.T.O. cum Programme Officer, Dist. Health and F.W. Samiti (RNTCP), Old Outdoor Complex, Suri, Birbhumb on deposition of non refundable fee of Rs. 200/- (Rupees Two Hundred only) in cash for each set of copy on all working days from 2.00 p.m. to 5.00 p.m. from 08.06.2009 to 12.06.2009 after showing Xerox copy of Trade License, sales tax Clearance Certificate, I.Tax Clearance Certificate, VAT Registration Certificate, & Drug License documents.

Tenders will be received by Registered Post only up to 4.30 p.m. on 19.06.2009 and will be opened at 3.00 p.m. on 23.06.2009.

Tender Committee reserves the right to accept or cancel bid without assigning any reasons whatsoever.

Terms and Conditions :-

- 1) The rate should be quoted against each items including all charges .
- 2) The Tenderer should be submitted Xerox copy with self attested Valid Trade License, sales tax Clearance Certificate, I.Tax Clearance Certificate, VAT Registration Certificate, & Drug License documents and M.S.D.S (Manufacturer Safety Data sheet).
- 3) The successful Tenderer shall have to supply the Lab. Consumable within the specified time , failing which the work order will be cancelled .
- 4) The Bid Security is Rs. 5000/- (Five thousand only) to be submitted in the form of Demand Draft from any Nationalized Bank drawn in favour of "District Health and Family Welfare Samity(RNTCP), Birbhumb".
- 5) The rates will remain valid for 1 (one) year from acceptance of Tender.



D.T.O cum Programme Officer
District H & F.W Samiti(RNTCP)
Birbhumb (W.B.)

Memo No.: DTO /Birbhumb/76/1(15)

Date:02/06/2009

Copy forwarded for information with request to display the same in the notice board:-

1. The Savadhipati , B.Z. Parishad cum President , Dist. H & F.W. Samiti, Birbhumb.
2. The District Magistrate cum Executive Vice-President, Dist. H & F.W. Samiti, Birbhumb.
3. The Addl. DHS (TB) Cum STO, Swasthya Bhavan, W.B. with request to display it in website (Dept. of Health).
4. The Addl. District Magistrate (D), Birbhumb
5. The Chief Medical Officer of Health , Birbhumb cum Secretary , Dist. H & F.W. Samiti, Birbhumb
- 6-10) Dy. CMOH-I/II/III/DMCHO/ZLO, Birbhumb
11. The Swasthya Karmadakshyay, BZP, Birbhumb
12. Officer in Charge , Health , Birbhumb
- 13)Accounts Officer, DH&FWS, Birbhumb
14. D.T.O., All W.B.
15. The Notice Board


D.T.O cum Programme Officer
District H & F.W Samiti(RNTCP)
Birbhumb (W.B.)


Specification of Lab. Consumable

Sl. No	Name of the Lab. Consumable	Specification	Approx Quantity
1	Plastic Sputum Containers	Cups made of Special Medical Grade Polypropylene, thin plastic, translucent, Diameter - 4cm, Capacity - 30 ml. Screwable Cap should also be made of and should be air tight. Leak	100000 Pcs
2	Slides for Microscope	Glass slides plain, Size = 76mm x 26mm x 1.3mm, clean, scratch free with smooth edges, uniform refractive index, pack of 50.	100000 Pcs
3	Wooden (Broom) Stick	Hard, Long, Straight & Fine.	120 Kg
4	Adhesive Labels for Sputum Containers	Soft Paper, Good Quality of Adhesive, Size:3" & 1"	110000 Pcs
5	Lens Paper	Soft microscope lens cleaning tissue, 4" x 6" booklet, each booklet containing 100 sheets.	120 Packs
6	White Absorbent Cotton (450 gms)	Absorbent, each roll of 450 Gm Colour: White	60 Roll
7	Filter Paper	Filter paper No. 1, diameter:- 12.5 cms for filtering carbol fuchsin using a small funnel of 10cm maximum diameter, Smooth on one side and rough on the other side, water permeable	130 Pack
8	Diamond Markers	6" (15.24 cm.) holder with artificial diamond (hard stone) embedded at one end with screw cap, to mark on microscope glass slides.	60 Pcs
9	Concentrate Sulphuric Acid 5 ltr Jar	H ₂ SO ₄ , molecular wt. 98.08, Minimum Assay: 98%, colour: clear	120 X 5 Lit
10	Basic Fuchsin - 25 gms	Chemical name :Pararosaniline hydrochloride, chemical structure: C ₂₀ H ₂₀ ClN ₃ mol wt:337.86Dye content: Approx.85%-88%(dye content must be mentioned), Colour: Metallic	180 X 25 gm
11	Methylene Blue - 25 gms	Chemical Name: Methylthionine chloride, Chemical Structure: C ₁₆ H ₁₈ ClN ₃ S, molecular Wt: 319.9, Dye content: approx.82% (should be available on the container).	15 X 25 gm
12	Immersion Oil - 30 ml	use for microscopy Refractive index(n _{23/D})-1.515 , Dencity (d _{20/4}): 1.303, Viscocity mPa.s-1500, Light transmittance passes free from polychlorobiphenyl (PCB)	150 X 30 ml
13	Phenol - 500 gms	Chemical name: Phenol, Chemical structure: C ₆ H ₅ OH, Molecular wt:94.11, Melting point:40 °C + 2, Purity:99.5%	50 X 500 gm
14	Methylated Spirit - 01 ltr	Chemical Name: Ethanol denatured + 5% Isopropyl alcohol + 5% Methanol, Molecular structure: C ₂ H ₅ OH, Molecular wt: 46.07, Purity: 90%.	150 Lit
15	Slide Boxes of (25 slides)	For storing 25 Slides, Plastic storage boxes, approximate dimensions:	90 Pcs
16	Slide Boxes of (100 slides)	For storing 100 Slides, Plastic storage boxes, approximate dimensions: 8.75" (22.2 cm) x 6.75"(17.15 cm) x 1.3"(3.3 cm)	60 pcs
17	Glass Rod (2.5' x 4 mm)	Glass, 2.5' x 4 mm	40 Pcs
18	Spirit Lamp (Steel)	Steel Made, Medium Size	60 Pcs
19	Poly-Funnel	leak proof, made of polypropylene. (45 mm Diameter)	10 Pcs
20	Foot Operated Bucket (12 ltr. Quantity)	leak proof, made of polypropylene. Capacity-12 Lit.	20 Pcs
21	Rolling Tissue Paper	white, Soft, non bedded and absorbent	120 Pcs
22	Poly - Dropping Bottle (100 ml)	100ml capacity built-in stem all polypropylene; leak proof screw cap made of polypropylene.	100 Pcs
23	Poly - Dropping Bottle (500 ml)	500ml capacity built-in stem all polypropylene; leak proof screw cap made of polypropylene.	10 Pcs
24	Poly-Wash Bottle (500 ml)	500ml capacity built-in stem all polypropylene; leak proof screw cap made of polypropylene.	50 Pcs
25	Duster	Cotton, Size: 2ft X 1.5 ft	60 Pcs
26	Hand Towel	Cotton, Size: 2ft X 1.5 ft	60 Pcs
27	Phenyl (40% Phenolic Compound)	Household disinfectant, containg phenolic compounds such as monochlorophenol, chloroxylenol, coal Tar Acid, Oils & Emulsifiers etc. The approximate content of phenolic	1200 Lit (For proper disposal as

			6 No.
	PH Paper	for 2 to 14 range	
	Disposal Syringe 5 ml	Disposable, Capacity: 5ml	15,000 Pcs
30	Distil water	5 ml, poly pack	15,000 pcs
31	Beacker (5 lit)	Glass ,autoclavable & heat resistant , capacity 5000ml.	6 Pcs
32	Container (2000 ml)	Glass stoppered bottles, 2000 ml capacity.	170 Pcs
33	Round Bottle Flask (5 lit)	Glass ware for stain preparation - Round (flat bottom) flask 5000 ml capacity.	12 Pcs
34	Analytical Balance	Weighing Capacity range: 1gm.- 1 kg, Readability 0.1gm. Pan size 6" (approx.) dual pans of Stainless Steel, mechanical, table top. Standard weights should be supplied along with the	2Pcs
35	Aluminum Pan	Dia: 9" / 10" capacity: 1.5 to 2 lit, good quality	6 Pcs
36	Conical Flask (250 ml)	Autoclavable and dry heat resistant, flat bottom, conical flask, Capacity-250 ml	20 Pcs
37	Electric Heater	1000 Watt	6 Pcs
38	Funnel	Clear Glass, Plain diameter: 5 cm	60 Pcs
39	Funnel	Clear Glass, Plain diameter: 10 cm	12 Pcs
40	Funnel	Clear Glass, Plain diameter: 18 cm	6 Pcs
41	Distilled Water Plant	Complete unit made in stainless steel., Constant level arrangement for inlet water, Wall hanging model, provided with MS Bracket for wall mounting, Boiling Chamber, Lid, Cooling	3 Pcs
42	Gloves (Disposable) 6.5"	Acid & alkali proof, Rubber, Anti-static, to cover at least 3/4th of forearm.	1000 pcs
43	Gloves (Rubber) 6.5"	Acid & alkali proof, Rubber, Anti-static, to cover at least 3/4th of forearm.	100 pcs
44	Face Musk	Cotton, Standard Size, Green colour	75 Pcs
45	Face Musk	Disposable, Standard Size	1000 Pcs
46	Apron (Cotton) Half Sleeve	Cotton, White colour, Standard Size.	70 Pcs

General requirements:-

1. All equipment quoted should comply with Indian Standards Institutions Guidelines or any other National or International Guidelines.
2. All the chemicals should have the Hazard Communication Standard from Chemical manufacturers and Material Safety Data Sheets (MSDS) for each chemical item should be provided to the purchaser of the product at the time of the order and thereafter, anytime the MSDS is significantly revised. Also, all the chemicals should have their detailed specifications, potency, date of manufacture, date of expiry, batch no. and precautions for use mentioned on the containers.


 District Tuberculosis Officer
 Birbhum